

Tillegg 6: Endringer, spørsmål og svar til konkurransen

Evt. endringer i konkurransegrunnlagets innhold vil markeres med rød skrift i de reviderte dokumentene.

Ref.	Spørsmål fra Tilbydere	Svar/endringer	Dato besvart/endret
6.1	Krav 3.3.13 Cabin Size definerer at "The cabin shall comply to EN13718-2 sub clause 4.10". Standard EN13718-2 sub clause 4.10 ekskluderer flere helikoptertyper som per i dag er i operasjon som ambulansehelikopter i Norge ved installasjon av en funksjonell EMS innredning, de oppfyller ikke minimum avstander så som beskrevet i standarden. Det bes om at Luftambulansetjenesten revurderer kravet eller tydeliggjør intensjonen med kravet så som det står i dag.	Oppdragiver har valg å besvare dette spørsmålet på engelsk (samme språk som standarden det henvises til). Dette for å unngå misforståelser. We encourage the bidder to read NS-EN 13718-2:2015 sub clause 4.10.1 and 4.10.2 closely. NS-EN 13718-2:2015 clause 4.10.1, defines the necessary functionality of the treatment area. Please note that the second sentence uses "should" as a verb. Clause 4.10.2 defines the recommended space needed for such functionality. As the second sentence in 4.10.2 state, the functionality (equivalent performance level) is mandatory, not the dimensions. The requirement Kravspesifikasjon Vedlegg A 3.3.13 remains unchanged.	1.februar 2016